

TO BE RESCINDED

4766-4-11

Maintenance requirements and documentation.

(A) Mobile intensive care unit vehicles

- (1) Provide a copy of the periodical maintenance plan for all vehicles that conform to manufacturer specifications;
- (2) All licensed organizations shall maintain documentation of all periodical maintenance and repairs performed on each vehicle to include:
 - (a) Date of service or repair;
 - (b) Description of service/repair performed;
 - (c) Documentation of who performed service/repair;
- (3) Out of service MoICU(s):
 - (a) The licensed organization shall place an out of service sign in a licensed MoICU(s) to be visible through the windshield with the following information included on the front of the sign:
 - (i) Date MoICU(s) was taken out of service;
 - (ii) Odometer reading at time MoICU(s) was taken out of service;
 - (iii) Reason MoICU(s) was taken out of service;
 - (iv) Printed name and signature of person responsible for taking vehicle out of service;
 - (v) The words "Out of Service" utilizing not less than three inch lettering.

(B) Bio-medical equipment

- (1) All licensed organizations shall maintain documentation of all periodical maintenance and repairs of bio-medical equipment as required by manufacturer and/or food and drug administration:

- (a) Date of last bio-medical inspection;
- (b) Date of service or repair;
- (c) Description of service/repair performed;
- (d) Documentation of who performed inspection, service, or repair.

Effective:

R.C. 119.032 review dates: 12/22/2011

Certification

Date

Promulgated Under: 119.03
Statutory Authority: 4766.03
Rule Amplifies: 4766.07
Prior Effective Dates: 04/18/2005